

CONSOLIDATED TEST RESULTS SUMMARY

Please see the following pages for full test results.

BULK SKU	BATCH #	LOQ: Limit Of Quantitation	
PRODUCT NAME	SERVING SIZE	LOD: Limit Of Detection	
LABORATORY :	OREGON ACCREDITATION: OR100028	1 g = 10 ⁻³ kg = 10 ³ mg = 10 ⁶ µg 1 mg/kg = 1 ppm = 1000 ppb	
POTENCY	PER SERVING	PER GRAM	Percent
Cannabidiol (CBD)	mg/serving	mg/g	%
Total THC (d9-THC, THCA)	mg/serving	mg/g	%
Cannabigerol (CBG)	mg/serving	mg/g	%
Cannabinol (CBN)	mg/serving	mg/g	%
Cannabichromene (CBC)	mg/serving	mg/g	%
Tetrahydrocannabinolic Acid (THCA)	mg/serving	mg/g	%
Delta-9-THC (d9-THC)	mg/serving	mg/g	%
Delta-8-THC (d8-THC)	mg/serving	mg/g	%
HEAVY METALS	PER SERVING	PER GRAM	REGULATORY ACTION LEVEL
Arsenic	µg/serving	µg/g	1.5 ppm
Cadmium	µg/serving	µg/g	0.5 ppm
Lead	µg/serving	µg/g	0.5 ppm
Mercury	µg/serving	µg/g	3.0 ppm
PESTICIDES	REGULATORY ACTION LEVEL		
None of the other 59 pesticides tested found above limit of detection in the sample.			10 ppb ^[1]
RESIDUAL SOLVENTS	Results	REGULATORY ACTION LEVEL	
Ethanol*	µg/g	5,000 ppm	
Heptane	µg/g	5,000 ppm	
None of the 34 residual solvents tested found above limit of quantitation in the sample.			
MICROBIAL	PASS/FAIL		
Yeast & Mold	Pass		
Coliform	Pass		



1. American Herbal Pharmacopoeia. (2014). Cannabis Inflorescence: Standards of Identity, Analysis, and Quality Control. Washington DC: AHP.

*Ethanol is a food additive used in some of our ingredients. The FDA has labeled ethanol as Generally Recognized as Safe (GRAS). Many foods contain trace amounts of ethanol, including soy sauce, pasta sauces, fruits and juices, etc. Our products contain safe levels of ethanol and always below pertinent regulatory action levels.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Customer: Etz Hayim Holdings
Product identity: FORM-GMY.RLX25-GA24
Client/Metric ID: .
Laboratory ID: 24-000754-0002

Summary

Potency:

Analyte per 1g	Result	Limits	Units	Status	
CBC per 1g	0.309		mg/1g		CBD-Total per Serving Size 5.01 mg/1g
CBD per 1g	5.01		mg/1g		
CBE per 1g	0.110		mg/1g		THC-Total per Serving Size 0.231 mg/1g
CBG per 1g	0.227		mg/1g		(Reported in milligrams per serving)
CBT per 1g	0.107		mg/1g		
Δ9-THC per 1g	0.231		mg/1g		

Residual Solvents:

Analyte	Result (µg/g)	Limits (µg/g)	Status
Ethanol	427		

Pesticides:

All analytes passing and less than LOQ.

Metals:

Less than LOQ for all analytes.

Microbiology:

Less than LOQ for all analytes.



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Customer: Etz Hayim Holdings
16427 NE Airport Way
PORTLAND 97230
United States of America (USA)

Product identity: FORM-GMY.RLX25-GA24

Client/Metric ID: .

Sample Date:

Laboratory ID: 24-000754-0002

Evidence of Cooling: No

Temp: 19.1 °C

Relinquished by: client

Serving Size #1: 1 g

Sample Results

Potency per 1g	Method: J AOAC 2015 V98-6 (mod) ^b	Units mg/se	Batch: 2400667	Analyze: 1/26/24 5:31:00 PM	
Analyte	Result	Limits	Units	LOQ	Notes
CBC per 1g	0.309		mg/1g	0.0316	
CBC-A per 1g	< LOQ		mg/1g	0.0316	
CBC-Total per 1g	0.309		mg/1g	0.0594	
CBD per 1g	5.01		mg/1g	0.0316	
CBD-A per 1g	< LOQ		mg/1g	0.0316	
CBD-Total per 1g	5.01		mg/1g	0.0594	
CBDV per 1g	< LOQ		mg/1g	0.0316	
CBDV-A per 1g	< LOQ		mg/1g	0.0316	
CBDV-Total per 1g	< LOQ		mg/1g	0.0591	
CBE per 1g	0.110		mg/1g	0.0316	
CBG per 1g	0.227		mg/1g	0.0316	
CBG-A per 1g	< LOQ		mg/1g	0.0316	
CBG-Total per 1g	0.227		mg/1g	0.0591	
CBL per 1g	< LOQ		mg/1g	0.0316	
CBL-A per 1g	< LOQ		mg/1g	0.0316	
CBL-Total per 1g	< LOQ		mg/1g	0.0594	
CBN per 1g	< LOQ		mg/1g	0.0316	
CBT per 1g	0.107		mg/1g	0.0316	
Δ8-THCV per 1g	< LOQ		mg/1g	0.0316	
Δ10-THC-9R per 1g	< LOQ		mg/1g	0.0316	
Δ10-THC-9S per 1g	< LOQ		mg/1g	0.0316	
Δ10-THC-Total per 1g	< LOQ		mg/1g	0.0633	
Δ8-THC per 1g	< LOQ		mg/1g	0.0316	
Δ9-THC per 1g	0.231		mg/1g	0.0316	
delta-9-THCP per 1g	< LOQ		mg/1g	0.0316	
exo-THC per 1g	< LOQ		mg/1g	0.0316	
THC-A per 1g	< LOQ		mg/1g	0.0316	
THC-Total per 1g	0.231		mg/1g	0.0594	
THCV per 1g	< LOQ		mg/1g	0.0316	
THCV-A per 1g	< LOQ		mg/1g	0.0316	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Potency per 1g						
Analyte	Method: J AOAC 2015 V98-6 (mod) ^P	Result	Limits	Units mg/se	Batch: 2400667	Analyze: 1/26/24 5:31:00 PM
THCV-Total per 1g		< LOQ		mg/1g		0.0594
Total Cannabinoids per 1g		5.99		mg/1g		

Microbiology							
Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status Notes
E.coli	< LOQ		cfu/g	10	2400565	01/27/24 AOAC 991.14 (Petrifilm) ^P	
Total Coliforms	< LOQ		cfu/g	10	2400565	01/27/24 AOAC 991.14 (Petrifilm) ^P	
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	2400566	01/28/24 AOAC 2014.05 (RAPID) ^P	
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	2400566	01/28/24 AOAC 2014.05 (RAPID) ^P	

Solvents											
Method: Residual Solvents by GC/MS ^P						Units µg/g		Batch 2400758		Analyze 01/31/24 12:36 PM	
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane)	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane)	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethanol	427		200		
Ethyl acetate	< LOQ	5000	200	pass		Ethyl benzene	< LOQ		200		
Ethyl ether	< LOQ	5000	200	pass		Ethylene glycol	< LOQ	620	200	pass	
Ethylene oxide	< LOQ	50.0	20.0	pass		Hexanes (sum)	< LOQ	290	150	pass	
Isopropyl acetate	< LOQ	5000	200	pass		Isopropylbenzene (Cumene)	< LOQ	70.0	30.0	pass	
m,p-Xylene	< LOQ		200			Methanol	< LOQ	3000	200	pass	
Methylene chloride	< LOQ	600	60.0	pass		Methylpropane (Isobutane)	< LOQ		200		
n-Butane	< LOQ		200			n-Heptane	< LOQ	5000	200	pass	
n-Hexane	< LOQ		30.0			n-Pentane	< LOQ		200		
o-Xylene	< LOQ		200			Pentanes (sum)	< LOQ	5000	600	pass	
Propane	< LOQ	5000	200	pass		Tetrahydrofuran	< LOQ	720	100	pass	
Toluene	< LOQ	890	100	pass		Total Xylenes	< LOQ		400		
Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass							



Pesticides											
Method: AOAC 2007.01 & EN 15662 (mod) ^b						Units mg/kg	Batch 2400594	Analyze 01/25/24 02:41 PM			
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Abamectin [‡]	< LOQ	0.50	0.250	pass		Acephate [‡]	< LOQ	0.40	0.200	pass	
Acequinocyl [‡]	< LOQ	2.0	1.00	pass		Acetamiprid [‡]	< LOQ	0.20	0.100	pass	
Aldicarb [‡]	< LOQ	0.40	0.200	pass		Azoxystrobin [‡]	< LOQ	0.20	0.100	pass	
Bifentazate [‡]	< LOQ	0.20	0.100	pass		Bifenthrin [‡]	< LOQ	0.20	0.100	pass	
Boscalid [‡]	< LOQ	0.40	0.200	pass		Carbaryl [‡]	< LOQ	0.20	0.100	pass	
Carbofuran [‡]	< LOQ	0.20	0.100	pass		Chlorantraniliprole [‡]	< LOQ	0.20	0.100	pass	
Chlorfenapyr [‡]	< LOQ	1.0	0.500	pass		Chlorpyrifos [‡]	< LOQ	0.20	0.100	pass	
Clofentezine [‡]	< LOQ	0.20	0.100	pass		Cyfluthrin [‡]	< LOQ	1.0	0.500	pass	
Cypermethrin [‡]	< LOQ	1.0	0.500	pass		Daminozide [‡]	< LOQ	1.0	0.500	pass	
Diazinon [‡]	< LOQ	0.20	0.100	pass		Dichlorvos [‡]	< LOQ	1.0	0.500	pass	
Dimethoate [‡]	< LOQ	0.20	0.100	pass		Ethoprophos [‡]	< LOQ	0.20	0.100	pass	
Etofenprox [‡]	< LOQ	0.40	0.200	pass		Etoazole [‡]	< LOQ	0.20	0.100	pass	
Fenoxycarb [‡]	< LOQ	0.20	0.100	pass		Fenpyroximate [‡]	< LOQ	0.40	0.200	pass	
Fipronil [‡]	< LOQ	0.40	0.200	pass		Flonicamid [‡]	< LOQ	1.0	0.400	pass	
Fludioxonil [‡]	< LOQ	0.40	0.200	pass		Hexythiazox [‡]	< LOQ	1.0	0.400	pass	
Imazalil [‡]	< LOQ	0.20	0.100	pass		Imidacloprid [‡]	< LOQ	0.40	0.200	pass	
Kresoxim-methyl [‡]	< LOQ	0.40	0.200	pass		Malathion [‡]	< LOQ	0.20	0.100	pass	
Metalaxyl [‡]	< LOQ	0.20	0.100	pass		Methiocarb [‡]	< LOQ	0.20	0.100	pass	
Methomyl [‡]	< LOQ	0.40	0.200	pass		MGK-264 [‡]	< LOQ	0.20	0.100	pass	
Myclobutanil [‡]	< LOQ	0.20	0.100	pass		Naled [‡]	< LOQ	0.50	0.250	pass	
Oxamyl [‡]	< LOQ	1.0	0.500	pass		Pacllobutrazole [‡]	< LOQ	0.40	0.200	pass	
Parathion-Methyl [‡]	< LOQ	0.20	0.100	pass		Permethrin [‡]	< LOQ	0.20	0.100	pass	
Phosmet [‡]	< LOQ	0.20	0.100	pass		Piperonyl butoxide [‡]	< LOQ	2.0	1.00	pass	
Prallethrin [‡]	< LOQ	0.20	0.100	pass		Propiconazole [‡]	< LOQ	0.40	0.200	pass	
Propoxur [‡]	< LOQ	0.20	0.100	pass		Pyrethrin I (total) [‡]	< LOQ	1.0	0.500	pass	
Pyridaben [‡]	< LOQ	0.20	0.100	pass		Spinosad [‡]	< LOQ	0.20	0.100	pass	
Spiromesifen [‡]	< LOQ	0.20	0.100	pass		Spirotetramat [‡]	< LOQ	0.20	0.100	pass	
Spiroxamine [‡]	< LOQ	0.40	0.200	pass		Tebuconazole [‡]	< LOQ	0.40	0.200	pass	
Thiacloprid [‡]	< LOQ	0.20	0.100	pass		Thiamethoxam [‡]	< LOQ	0.20	0.100	pass	
Trifloxystrobin [‡]	< LOQ	0.20	0.100	pass							

Metals										
Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method		Status	Notes	
Arsenic [‡]	< LOQ	0.200	mg/kg	0.0157	2400704	01/29/24	AOAC 2013.06 (mod.) ^b	pass		
Cadmium [‡]	< LOQ	0.200	mg/kg	0.0157	2400704	01/29/24	AOAC 2013.06 (mod.) ^b	pass		
Lead [‡]	< LOQ	0.500	mg/kg	0.0157	2400704	01/29/24	AOAC 2013.06 (mod.) ^b	pass		
Mercury [‡]	< LOQ	0.100	mg/kg	0.00783	2400704	01/29/24	AOAC 2013.06 (mod.) ^b	pass		



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

Ⓟ = ISO/IEC 17025:2017 accredited method.

Ⓜ = TNI accredited analyte.

Units of Measure

cfu/g = Colony forming units per gram

g = g

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/1g = Milligram per 1g

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner
General Manager



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Revision: 3 Document ID: 3120
 Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC2007.1 &EN 15662		Units: mg/Kg			Batch ID 2400594			
Method Bank	Laboratory Control Sample							
Analyte	Blank Result	Blank Limits	Notes	LCS Result	LCS Spke	LCS% Re	Limits	Notes
Abamectin	0.000	< 0.250		0.891	1.000	89.1	50.0	150
Acephate	0.153	< 0.200		0.645	0.800	80.6	60.0	120
Acetaminocyl	0.000	< 1.000		3.690	4.000	92.2	40.0	160
Acetamiprid	0.000	< 0.100		0.343	0.400	85.8	60.0	120
Aldicarb	0.000	< 0.200		0.835	0.800	104.4	60.0	120
Azoxystrobin	0.000	< 0.100		0.358	0.400	89.6	60.0	120
Bifenazate	0.000	< 0.100		0.375	0.400	93.8	60.0	120
Bifenthrin	0.000	< 0.100		0.389	0.400	97.2	50.0	150
Boscalid	0.000	< 0.200		0.781	0.800	97.6	60.0	120
Carbaryl	0.000	< 0.100		0.359	0.400	89.9	60.0	120
Carbofuran	0.000	< 0.100		0.346	0.400	86.5	60.0	120
Chlorantraniliprole	0.000	< 0.100		0.385	0.400	96.3	60.0	120
Chlorfenapyr	0.000	< 0.500		1.985	2.000	99.3	60.0	120
Chlorpyrifos	0.000	< 0.100		0.347	0.400	86.8	60.0	120
Clofentazine	0.000	< 0.100		0.347	0.400	86.7	60.0	120
Cyfluthrin	0.000	< 0.500		1.872	2.000	93.6	50.0	150
Cypermethrin	0.000	< 0.500		1.859	2.000	92.9	50.0	150
Daminozide	0.000	< 0.500		0.687	2.000	34.4	60.0	120
Diazinon	0.000	< 0.100		0.403	0.400	100.7	60.0	120
Dichlorvos	0.000	< 0.500		1.787	2.000	89.4	60.0	120
Dimethoate	0.000	< 0.100		0.339	0.400	84.8	60.0	120
Ethoprophos	0.000	< 0.100		0.383	0.400	95.7	60.0	120
Etofenprox	0.000	< 0.200		0.749	0.800	93.6	50.0	150
Etoxazole	0.000	< 0.100		0.375	0.400	93.7	60.0	120
Fenoxycarb	0.000	< 0.100		0.381	0.400	95.2	60.0	120
Fenpyroximate	0.000	< 0.200		0.734	0.800	91.7	60.0	120
Fipronil	0.000	< 0.200		0.774	0.800	96.7	60.0	120
Fonicamid	0.000	< 0.250		0.870	1.000	87.0	60.0	120
Fludioxonil	0.000	< 0.200		0.799	0.800	99.9	50.0	150
Hexythiazox	0.000	< 0.250		0.921	1.000	92.1	60.0	120
Imazalil	0.000	< 0.100		0.353	0.400	88.2	60.0	120
Imidacloprid	0.000	< 0.200		0.699	0.800	87.4	60.0	120
Kresoxim-methyl	0.000	< 0.200		0.750	0.800	93.8	60.0	120
Malathion	0.000	< 0.100		0.388	0.400	96.9	60.0	120
Metaxalyl	0.000	< 0.100		0.366	0.400	91.4	60.0	120
Methiocarb	0.000	< 0.100		0.348	0.400	87.0	60.0	120
Methomyl	0.000	< 0.200		0.695	0.800	86.9	60.0	120
MGK-264	0.000	< 0.100		0.386	0.400	96.5	50.0	150
Myclobutanil	0.000	< 0.100		0.355	0.400	88.7	60.0	120
Naled	0.000	< 0.250		0.848	1.000	84.8	50.0	150
Oxamyl	0.000	< 0.500		1.740	2.000	87.0	60.0	120
Pacllobutrazole	0.000	< 0.200		0.753	0.800	94.1	60.0	120
Parathion-Methyl	0.006	< 0.100		0.388	0.400	97.0	50.0	150
Permethrin	0.000	< 0.100		0.360	0.400	90.1	50.0	150
Phosmet	0.000	< 0.100		0.371	0.400	92.8	50.0	150
Piperonyl butoxide	0.000	< 0.500		1.907	2.000	95.3	60.0	120
Prallethrin	0.000	< 0.100		0.368	0.400	91.9	60.0	120
Propiconazole	0.000	< 0.200		0.596	0.800	74.5	60.0	120
Propoxur	0.000	< 0.100		0.422	0.400	105.5	60.0	120
Pyrethrin (Summe)	0.001	< 0.100		0.449	0.488	92.0	60.0	120
Pyridaben	0.000	< 0.100		0.386	0.400	96.5	50.0	150
Spinosad	0.000	< 0.100		0.369	0.388	95.1	50.0	150
Spiromesifen	0.000	< 0.100		0.368	0.400	92.1	60.0	120
Spirotetramat	0.000	< 0.100		0.390	0.400	97.5	60.0	120
Spiroxamine	0.000	< 0.200		0.596	0.800	74.5	60.0	120
Tebuconazole	0.000	< 0.200		0.746	0.800	93.3	60.0	120
Thiacloprid	0.000	< 0.100		0.354	0.400	88.4	60.0	120
Thiamethoxam	0.000	< 0.100		0.303	0.400	75.8	60.0	120
Trifloxystrobin	0.000	< 0.100		0.368	0.400	92.0	60.0	120

Q7



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Revision: 3 Document ID: 3120
Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC2007.1 & EN 15662		Units: mg/Kg				Batch ID 2400594				
Matrix Spke/Matrix Spke Duplicate Recoveries	Result	MS Res	MSD Res	Spike	RFD%	Limit	MS % Re	MSD % Re	Limits	Notes
Abamectin	0.000	0.999	1.019	1.000	1.9%	< 30	99.9%	101.9%	50 - 150	
Acephate	0.000	0.779	0.896	0.800	13.9%	< 30	97.4%	111.9%	50 - 150	
Acetaminocyl	0.000	4.043	4.121	4.000	1.9%	< 30	101.1%	103.0%	50 - 150	
Acetamidprid	0.000	0.428	0.434	0.400	1.5%	< 30	106.9%	108.5%	50 - 150	
Aldicarb	0.000	0.888	0.831	0.800	6.6%	< 30	111.0%	103.9%	50 - 150	
Azoxystrobin	0.000	0.384	0.364	0.400	5.3%	< 30	96.1%	91.1%	50 - 150	
Bifenazate	0.000	0.404	0.396	0.400	2.2%	< 30	101.1%	98.9%	50 - 150	
Bifenthrin	0.000	0.434	0.468	0.400	7.5%	< 30	108.6%	117.0%	50 - 150	
Boscalid	0.000	0.797	0.784	0.800	1.7%	< 30	99.6%	98.0%	50 - 150	
Carbaryl	0.000	0.385	0.399	0.400	3.4%	< 30	96.3%	99.7%	50 - 150	
Carbofuran	0.000	0.379	0.396	0.400	4.3%	< 30	94.8%	98.9%	50 - 150	
Chlorantraniliprole	0.000	0.392	0.421	0.400	7.3%	< 30	97.9%	105.3%	50 - 150	
Chlorfenapyr	0.000	2.245	2.383	2.000	6.0%	< 30	112.2%	119.1%	50 - 150	
Chlorpyrifos	0.000	0.363	0.375	0.400	3.4%	< 30	90.7%	93.8%	50 - 150	
Clofentezine	0.000	0.232	0.232	0.400	0.1%	< 30	58.0%	58.1%	50 - 150	
Cyfluthrin	0.000	2.348	2.466	2.000	4.9%	< 30	117.4%	123.3%	30 - 150	
Cypermethrin	0.000	2.513	2.570	2.000	2.3%	< 30	125.7%	128.5%	50 - 150	
Daminozide	0.000	0.834	0.864	2.000	3.5%	< 30	41.7%	43.2%	30 - 150	
Diazinon	0.000	0.409	0.401	0.400	1.9%	< 30	102.3%	100.3%	50 - 150	
Dichlorvos	0.000	2.039	2.026	2.000	0.6%	< 30	101.9%	101.3%	50 - 150	
Dimethoate	0.000	0.372	0.404	0.400	8.2%	< 30	93.0%	101.0%	50 - 150	
Ethoprophos	0.000	0.387	0.386	0.400	0.2%	< 30	96.7%	96.5%	50 - 150	
Etofenprox	0.000	0.763	0.825	0.800	7.8%	< 30	95.4%	103.2%	50 - 150	
Etoxazole	0.000	0.382	0.409	0.400	6.9%	< 30	95.5%	102.4%	50 - 150	
Fenoxycarb	0.000	0.394	0.400	0.400	1.5%	< 30	98.5%	100.0%	50 - 150	
Fenpyroximate	0.000	0.776	0.836	0.800	7.4%	< 30	97.0%	104.5%	50 - 150	
Fipronil	0.000	0.845	0.859	0.800	1.6%	< 30	105.7%	107.4%	50 - 150	
Fonicamid	0.000	1.055	1.140	1.000	7.7%	< 30	105.5%	114.0%	50 - 150	
Fludioxonil	0.000	0.797	0.783	0.800	1.9%	< 30	99.7%	97.8%	50 - 150	
Hexythiazox	0.000	2.165	2.304	1.000	6.2%	< 30	216.5%	230.4%	50 - 150	Q
Imazalil	0.000	0.401	0.402	0.400	0.2%	< 30	100.2%	100.4%	50 - 150	
Imidacloprid	0.000	0.826	0.864	0.800	4.4%	< 30	103.3%	107.9%	50 - 150	
Kresoxim-methyl	0.000	0.820	0.769	0.800	6.3%	< 30	102.5%	96.2%	50 - 150	
Malathion	0.000	0.386	0.402	0.400	4.2%	< 30	96.4%	100.5%	50 - 150	
Metaxalyl	0.000	0.381	0.394	0.400	3.4%	< 30	95.2%	98.5%	50 - 150	
Methiocarb	0.000	0.413	0.421	0.400	1.9%	< 30	103.3%	105.3%	50 - 150	
Methomyl	0.000	0.776	1.001	0.800	25.3%	< 30	97.0%	125.1%	50 - 150	
MGK-264	0.000	0.391	0.435	0.400	10.6%	< 30	97.8%	108.7%	50 - 150	
Myclobutanil	0.000	0.403	0.387	0.400	4.0%	< 30	100.7%	96.7%	50 - 150	
Naled	0.000	0.869	0.927	1.000	6.5%	< 30	86.9%	92.7%	50 - 150	
Oxamyl	0.000	1.917	1.871	2.000	2.5%	< 30	95.9%	93.5%	50 - 150	
Pacllobutrazole	0.000	0.829	0.805	0.800	2.9%	< 30	103.6%	100.6%	50 - 150	
Parathion-Methyl	0.002	0.391	0.423	0.400	8.0%	< 30	97.4%	105.4%	30 - 150	
Permethrin	0.000	0.408	0.423	0.400	3.4%	< 30	102.1%	105.6%	50 - 150	
Phosmet	0.000	0.414	0.425	0.400	2.5%	< 30	103.5%	106.2%	50 - 150	
Piperonyl butoxide	0.000	1.832	1.884	2.000	2.8%	< 30	91.6%	94.2%	50 - 150	
Prallethrin	0.000	0.411	0.409	0.400	0.6%	< 30	102.8%	102.3%	50 - 150	
Propiconazole	0.000	0.607	0.633	0.800	4.2%	< 30	75.9%	79.1%	50 - 150	
Propoxur	0.000	0.458	0.497	0.400	8.1%	< 30	114.5%	124.1%	50 - 150	
Pyrethrin (Summe)	0.000	0.469	0.478	0.488	2.1%	< 30	96.0%	98.0%	50 - 150	
Pyridaben	0.000	0.468	0.467	0.400	0.1%	< 30	116.9%	116.8%	50 - 150	
Spinosad	0.000	0.398	0.402	0.388	1.1%	< 30	102.5%	103.6%	50 - 150	
Spiromesifen	0.000	0.380	0.390	0.400	2.6%	< 30	95.0%	97.5%	50 - 150	
Spirotetramat	0.000	0.428	0.422	0.400	1.4%	< 30	107.1%	105.6%	50 - 150	
Spiroxamine	0.000	0.615	0.634	0.800	3.0%	< 30	76.9%	79.2%	50 - 150	
Tebuconazole	0.000	0.759	0.761	0.800	0.3%	< 30	94.8%	95.1%	50 - 150	
Thiacloprid	0.000	0.396	0.407	0.400	2.6%	< 30	99.1%	101.7%	50 - 150	
Thiamethoxam	0.000	0.337	0.407	0.400	18.7%	< 30	84.3%	101.7%	50 - 150	
Trifloxystrobin	0.000	0.386	0.393	0.400	2.0%	< 30	96.4%	98.3%	50 - 150	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Revision: 4 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 Batch ID: 2400667

Laboratory Control Sample									
Analyte	LCS	Result	Spike	Units	% Rec	Limits		Evaluation	Notes
CBDVA	2	0.0316	0.0324	%	97.6	80.0	- 120	Acceptable	
CBDV	2	0.0332	0.0347	%	95.6	80.0	- 120	Acceptable	
CBE	2	0.0337	0.0350	%	96.3	80.0	- 120	Acceptable	
CBDA	1	0.0295	0.0306	%	96.2	90.0	- 110	Acceptable	
CBGA	1	0.0293	0.0303	%	96.5	80.0	- 120	Acceptable	
CBG	1	0.0311	0.0327	%	95.1	80.0	- 120	Acceptable	
CBD	1	0.0319	0.0335	%	95.1	90.0	- 110	Acceptable	
THCV	2	0.0336	0.0345	%	97.5	80.0	- 120	Acceptable	
d8THCV	2	0.0289	0.0303	%	95.3	80.0	- 120	Acceptable	
THCVA	2	0.0311	0.0318	%	97.7	80.0	- 120	Acceptable	
CBN	1	0.0288	0.0301	%	95.4	80.0	- 120	Acceptable	
exo-THC	2	0.0315	0.0330	%	95.3	80.0	- 120	Acceptable	
d9THC	1	0.0301	0.0317	%	95.1	90.0	- 110	Acceptable	
d8THC	1	0.0290	0.0304	%	95.1	90.0	- 110	Acceptable	
9S-d10THC	1	0.0306	0.0324	%	94.4	80.0	- 120	Acceptable	
CBL	2	0.0303	0.0326	%	93.0	80.0	- 120	Acceptable	
9R-d10THC	1	0.0295	0.0317	%	93.1	80.0	- 120	Acceptable	
CBC	2	0.0331	0.0349	%	95.0	80.0	- 120	Acceptable	
THCA	1	0.0289	0.0296	%	97.4	90.0	- 110	Acceptable	
CBCA	2	0.0324	0.0334	%	96.7	80.0	- 120	Acceptable	
CBLA	2	0.0328	0.0338	%	96.8	80.0	- 120	Acceptable	
d9THCP	2	0.0305	0.0328	%	93.2	80.0	- 120	Acceptable	
CBT	2	0.0317	0.0347	%	91.5	80.0	- 120	Acceptable	

Method Blank						
Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBDV	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBE	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBDA	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBGA	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBG	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBD	<LOQ	0.00318	%	< 0.00318	Acceptable	
THCV	<LOQ	0.00318	%	< 0.00318	Acceptable	
d8THCV	<LOQ	0.00318	%	< 0.00318	Acceptable	
THCVA	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBN	<LOQ	0.00318	%	< 0.00318	Acceptable	
exo-THC	<LOQ	0.00318	%	< 0.00318	Acceptable	
d9THC	<LOQ	0.00318	%	< 0.00318	Acceptable	
d8THC	<LOQ	0.00318	%	< 0.00318	Acceptable	
9S-d10THC	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBL	<LOQ	0.00318	%	< 0.00318	Acceptable	
9R-d10THC	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBC	<LOQ	0.00318	%	< 0.00318	Acceptable	
THCA	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBCA	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBLA	<LOQ	0.00318	%	< 0.00318	Acceptable	
d9THCP	<LOQ	0.00318	%	< 0.00318	Acceptable	
CBT	<LOQ	0.00318	%	< 0.00318	Acceptable	

Abbreviations
 ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:
 % - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Revision: 4 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6		Batch ID: 2400667						
Sample Duplicate		Sample ID: 24-000702-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBDV	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBD	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBDA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBGA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBG	0.0172	0.0169	0.00330	%	1.29	< 20	Acceptable	
CBD	0.601	0.595	0.00330	%	1.02	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
d8THCV	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBN	0.115	0.114	0.00330	%	1.04	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
d9THC	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
d8THC	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
9S-d10THC	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
9R-d10THC	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBC	0.00509	0.00508	0.00330	%	0.128	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBCA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
d9THCP	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	
CBT	<LOQ	<LOQ	0.00330	%	NA	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:

% - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794

Report Number: 24-000754/D003.R000
 Report Date: 01/31/2024
 ORELAP#: OR100028
 Purchase Order: 2824295
 Received: 01/24/24 16:13



Revision: 2 Document ID: 7087
 Legacy ID: CFL-E33Effective:

Laboratory Quality Control Results

Residual Solvents				Batch ID: 2400758					
Method Blank				Laboratory Control Sample					
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
Propane	ND	< 200		617	584	µg/g	105.7	60 - 120	
Isobutane	ND	< 200		662	767	µg/g	86.3	60 - 120	
Butane	ND	< 200		698	782	µg/g	89.3	60 - 120	
2,2-Dimethylpropane	ND	< 200		1080	939	µg/g	115.0	60 - 120	
Methanol	ND	< 200		1490	1600	µg/g	93.1	60 - 120	
Ethylene Oxide	ND	< 30		59.2	57.1	µg/g	103.7	60 - 120	
2-Methylbutane	ND	< 200		1470	1600	µg/g	91.9	60 - 120	
Pentane	ND	< 200		1470	1600	µg/g	91.9	60 - 120	
Ethanol	ND	< 200		1420	1600	µg/g	88.8	70 - 130	
Ethyl Ether	ND	< 200		1500	1600	µg/g	93.8	60 - 120	
2,2-Dimethylbutane	ND	< 30		148	161	µg/g	91.9	60 - 120	
Acetone	ND	< 200		1510	1600	µg/g	94.4	60 - 120	
2-Propanol	ND	< 200		1540	1600	µg/g	96.3	60 - 120	
Ethyl Formate	ND	< 500		1380	1600	µg/g	86.3	70 - 130	
Acetonitrile	ND	< 100		460	488	µg/g	94.3	60 - 120	
Methyl Acetate	ND	< 500		1530	1610	µg/g	95.0	70 - 130	
2,3-Dimethylbutane	ND	< 30		149	163	µg/g	91.4	60 - 120	
Dichloromethane	ND	< 60		438	488	µg/g	89.8	60 - 120	
2-Methylpentane	ND	< 30		137	161	µg/g	85.1	60 - 120	
MTBE	ND	< 500		1650	1650	µg/g	100.0	70 - 130	
3-Methylpentane	ND	< 30		151	162	µg/g	93.2	60 - 120	
Hexane	ND	< 30		131	161	µg/g	81.4	60 - 120	
1-Propanol	ND	< 500		1640	1620	µg/g	101.2	70 - 130	
Methylethylketone	ND	< 500		1560	1610	µg/g	96.9	70 - 130	
Ethyl acetate	ND	< 200		1550	1610	µg/g	96.3	60 - 120	
2-Butanol	ND	< 200		1560	1610	µg/g	96.9	60 - 120	
Tetrahydrofuran	ND	< 100		439	483	µg/g	90.9	60 - 120	
Cyclohexane	ND	< 200		1430	1600	µg/g	89.4	60 - 120	
2-methyl-1-propanol	ND	< 500		1570	1600	µg/g	98.1	70 - 130	
Benzene	ND	< 1		2.83	4.99	µg/g	56.7	60 - 120	Q6
Isopropyl Acetate	ND	< 200		1520	1600	µg/g	95.0	60 - 120	
Heptane	ND	< 200		1510	1600	µg/g	94.4	60 - 120	
1-Butanol	ND	< 500		1560	1610	µg/g	96.9	70 - 130	
Propyl Acetate	ND	< 500		1620	1610	µg/g	100.6	70 - 130	
1,4-Dioxane	ND	< 100		392	480	µg/g	81.7	60 - 120	
2-Ethoxyethanol	ND	< 30		146	161	µg/g	90.7	60 - 120	
Methylisobutylketone	ND	< 500		1590	1610	µg/g	98.8	70 - 130	
3-Methyl-1-butanol	ND	< 500		1510	1610	µg/g	93.8	70 - 130	
Ethylene Glycol	ND	< 200		479	481	µg/g	99.6	60 - 120	
Toluene	ND	< 100		443	483	µg/g	91.7	60 - 120	
Isobutyl Acetate	ND	< 500		1610	1610	µg/g	100.0	70 - 130	
1-Pentanol	ND	< 500		1530	1610	µg/g	95.0	70 - 130	
Butyl Acetate	ND	< 500		1560	1600	µg/g	97.5	70 - 130	
Ethylbenzene	ND	< 200		862	962	µg/g	89.6	60 - 120	
m,p-Xylene	ND	< 200		865	972	µg/g	89.0	60 - 120	
o-Xylene	ND	< 200		864	965	µg/g	89.5	60 - 120	
Cumene	ND	< 30		142	169	µg/g	84.0	60 - 120	
Anisole	ND	< 500		1350	1600	µg/g	84.4	70 - 130	
DMSO	ND	< 500		1070	1600	µg/g	66.9	70 - 130	Q6
1,2-dimethoxyethane	ND	< 50		173	163	µg/g	106.1	70 - 130	
Triethylamine	ND	< 500		1180	1600	µg/g	73.8	70 - 130	
N,N-dimethylformamide	ND	< 150		424	482	µg/g	88.0	70 - 130	
N,N-dimethylacetamide	ND	< 150		374	483	µg/g	77.4	70 - 130	
Pyridine	ND	< 50		134	161	µg/g	83.2	70 - 130	
Sulfolane	ND	< 50		114	163	µg/g	69.9	70 - 130	
1,2-Dichloroethane	ND	< 1		0.932	1	µg/g	93.2	70 - 130	
Chloroform	ND	< 1		0.965	1	µg/g	96.5	70 - 130	
Trichloroethylene	ND	< 1		0.781	1	µg/g	78.1	70 - 130	
1,1-Dichloroethane	ND	< 1		0.897	1	µg/g	89.7	70 - 130	



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13

Revision: 2 Document ID: 7087
 Legacy ID: CFL-E33Effective:

QC - Sample Duplicate		Sample ID: 24-000580-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Formate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Methyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	60	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
1-Propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methyl ethyl ketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-methyl-1-propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
1-Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methyl-1-butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1-Pentanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Butyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,2-dimethoxyethane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
Pyridine	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Sulfolane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Trichloroethylene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
1,1-Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:

µg/g - Microgram per gram or ppm



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 24-000754/D003.R000
Report Date: 01/31/2024
ORELAP#: OR100028
Purchase Order: 2824295
Received: 01/24/24 16:13





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.